

INTERNATIONAL STANDARD

ISO
8518

Second edition
2001-12-15

Workplace air — Determination of particulate lead and lead compounds — Flame or electrothermal atomic absorption spectrometric method

*Air des lieux de travail — Dosage du plomb particulaire et des composés
particulaires du plomb — Méthode par spectrométrie d'absorption atomique
dans la flamme ou méthode par spectrométrie d'absorption avec
atomisation électrothermique*

STANDARDSISO.COM : Click to view the full PDF of ISO 8518:2001



Reference number
ISO 8518:2001(E)

© ISO 2001

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 8518:2001

© ISO 2001

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.ch
Web www.iso.ch

Printed in Switzerland

Contents

	Page
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Principle	5
5 Reactions	6
6 Requirement	6
7 Reagents	6
8 Apparatus	7
9 Occupational exposure assessment	11
9.1 Assessment strategy	11
9.2 Measurement strategy	11
9.3 Selection of measurement conditions and measurement pattern	11
10 Sampling	12
10.1 Preliminary considerations	12
10.2 Preparation of sampling equipment	13
10.3 Sampling position	14
10.4 Collection of samples	14
10.5 Transportation	14
11 Analysis	15
11.1 Cleaning of glassware and plasticware	15
11.2 Preparation of sample and blank solutions	15
11.3 Instrumental analysis	17
11.4 Estimation of the instrumental detection limit	20
11.5 Estimation of the method detection limit	20
11.6 Quality control	20
12 Expression of results	21
12.1 Calculation	21
12.2 Method performance	22
13 Special cases	23
14 Test report	23

Annexes

A Guidance on filter selection.....	25
B Temperature and pressure correction	27
Bibliography.....	29

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 8518 was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 2, *Workplace atmospheres*.

This second edition cancels and replaces the first edition (ISO 8518:1990), which has been technically revised.

Annexes A and B of this International Standard are for information only.

Introduction

The health of workers in many industries, e.g. mining, metal refining, battery manufacture, construction, etc., is at risk through exposure by inhalation of particulate lead and lead compounds. Industrial hygienists and other public health professionals need to determine the effectiveness of measures taken to control workers' exposure, and this is generally achieved by making workplace air measurements. This International Standard provides a method for making valid exposure measurements for lead. It will be of benefit to: agencies concerned with health and safety at work; industrial hygienists and other public health professionals; analytical laboratories; industrial users and workers of metals and metalloids, etc.

It has been assumed in the drafting of this International Standard that the execution of its provisions, and the interpretation of the results obtained, is entrusted to appropriately qualified and experienced people.

STANDARDSISO.COM : Click to view the full PDF of ISO 8518:2001

STANDARDSISO.COM : Click to view the full PDF of ISO 8518:2001

Workplace air — Determination of particulate lead and lead compounds — Flame or electrothermal atomic absorption spectrometric method

1 Scope

1.1 This International Standard specifies flame and electrothermal atomic absorption spectrometric methods for the determination of the time-weighted average mass concentration of particulate lead and lead compounds in workplace air.

1.2 The method is applicable to personal sampling of the inhalable fraction of airborne particles, as defined in ISO 7708, and to static (area) sampling.

1.3 The sample dissolution procedure specifies hot plate or microwave digestion, or ultrasonic extraction (11.2). The use of an alternative, more vigorous dissolution procedure is necessary when it is desired to extract lead from compounds present in the test atmosphere that are insoluble using the dissolution procedures described herein.

1.4 The flame atomic absorption method is applicable to the determination of masses of approximately 1 µg to 200 µg of lead per sample, without dilution [1]. The electrothermal atomic absorption method is applicable to the determination of masses of approximately 0,01 µg to 0,5 µg of lead per sample, without dilution [1].

1.5 The ultrasonic extraction procedure has been validated for the determination of masses of approximately 20 µg to 100 µg of lead per sample, for laboratory-generated lead fume air filter samples [2].

1.6 The concentration range for lead in air for which this procedure is applicable is determined in part by the sampling procedure selected by the user (see 10.1).

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 648:1977, *Laboratory glassware — One-mark pipettes*

ISO 1042:1998, *Laboratory glassware — One-mark volumetric flasks*

ISO 3585:1998, *Borosilicate glass 3.3 — Properties*

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ISO 7708:1995, *Air quality — Particle size fraction definitions for health-related sampling*

ISO 8655-1, *Piston-operated volumetric apparatus — Part 1: Terminology, general requirements and user recommendations*

ISO 8655-2, *Piston-operated volumetric apparatus — Part 2: Piston pipettes*

ISO 8655-5, *Piston-operated volumetric apparatus — Part 5: Dispensers*

ISO 8655-6, *Piston-operated volumetric apparatus — Part 6: Gravimetric methods for the determination of measurement error*

ISO 15202-2:2001, *Workplace air — Determination of metals and metalloids in airborne particulate matter by inductively coupled plasma atomic emission spectrometry — Part 2: Sample preparation*

EN 13205¹⁾, *Workplace atmospheres — Assessment of performance of instruments for measurement of airborne particle concentrations*

3 Terms and definitions

For the purposes of this International Standard, the following terms and definitions apply.

3.1 General definitions

3.1.1

chemical agent

any chemical element or compound, on its own or admixed as it occurs in the natural state or as produced, used or released, including release as waste, by any work activity, whether or not produced intentionally and whether or not placed on the market

[EN 1540]

3.1.2

breathing zone

space around a person's face from where he or she takes his or her breath

NOTE For technical purposes, a more precise definition is as follows: hemisphere (generally accepted to be 0,3 m in radius) extending in front of the human face, centred on the midpoint of a line joining the ears; the base of the hemisphere is a plane through this line, the top of the head and the larynx. This definition is not applicable when respiratory protective equipment is used.

[EN 1540]

3.1.3

exposure by inhalation

situation in which a chemical agent is present in air which is inhaled by a person

3.1.4

measuring procedure

procedure for sampling and analysing one or more chemical agents in the air, including storage and transportation of the sample

3.1.5

operating time

period during which a sampling pump can be operated at specified flowrate and back-pressure without recharging or replacing the battery

[EN 1232]

3.1.6

time-weighted average concentration

TWA concentration

concentration of a chemical agent in the atmosphere, averaged over the reference period

1) To be published.

NOTE A more detailed discussion of TWA concentrations and their use can be found in [3].

3.1.7

limit value

reference figure for concentration of a chemical agent in air

EXAMPLE Threshold Limit Value[®] (TLV) for a given substance in workplace air, as established by the ACGIH [3].

3.1.8

reference period

specified period of time stated for the limit value of a specific chemical agent

NOTE Examples of limit values for different reference periods are short-term and long-term exposure limits, such as those established by the ACGIH [3].

3.1.9

workplace

defined area or areas in which work activities are carried out

[EN 1540]

3.2 Particle size fraction definitions

3.2.1

inhalable convention

target specification for sampling instruments when the inhalable fraction is of interest

[ISO 7708]

3.2.2

inhalable fraction

mass fraction of total airborne particles which is inhaled through the nose and mouth

NOTE The inhalable fraction depends on the speed and direction of air movement, on breathing rate and other factors.

[ISO 7708]

3.2.3

respirable convention

target specification for sampling instruments when the respirable fraction is of interest

[ISO 7708]

3.2.4

respirable fraction

mass fraction of inhaled particles penetrating to the unciliated airways

[ISO 7708]

3.2.5

total airborne particles

all particles surrounded by air in a given volume of air

NOTE Because all measuring instruments are size-selective to some extent, it is often impossible to measure the total concentration of airborne particles.

[ISO 7708]

3.3 Sampling definitions

3.3.1

personal sampler

device attached to a person that samples air in the breathing zone

[EN 1540]

3.3.2

personal sampling

process of sampling carried out using a personal sampler

[EN 1540]

3.3.3

sampling instrument

sampler

device for collecting airborne particles

NOTE Instruments used to collect airborne particles are frequently referred to by a number of other terms, e.g. sampling heads, filter holders, filter cassettes, etc.

3.3.4

static sampling

area sampling

process of air sampling carried out in a particular location

3.3.5

static sampler

area sampler

device, not attached to a person, used in static sampling

3.4 Definitions used in analysis

3.4.1

sample dissolution

process of obtaining a solution containing the analytes of interest from a sample

NOTE This may or may not involve complete dissolution of the sample.

3.4.2

sample preparation

all operations carried out on a sample, after transportation and storage, to prepare it for analysis, including transformation of the sample into a measurable state, where necessary

3.4.3

sample solution

solution prepared by the process of sample dissolution, but possibly needing to be subjected to further operations in order to produce a test solution that is ready for analysis

3.4.4

test solution

solution prepared by the process of sample dissolution and, if necessary, having been subjected to any further operations required to bring it into a state in which it is ready for analysis

3.5 Statistical terms

3.5.1

analytical recovery

ratio of the mass of analyte measured when a sample is analysed to the known mass of analyte in that sample

NOTE It is expressed as a percentage.

3.5.2**bias**

consistent deviation of the results of a measurement process from the true value of the air quality characteristic itself

[ISO 6879]

3.5.3**overall uncertainty**

⟨of a measuring procedure or of an instrument⟩ quantity used to characterize as a whole the uncertainty of a result given by an apparatus or measuring procedure

NOTE It is expressed, as a percentage, by a combination of bias and precision, usually according to the formula:

$$\frac{|\bar{x} - x_{\text{ref}}| + 2s}{x_{\text{ref}}} \times 100$$

where

\bar{x} is the mean value of results of a number (n) of repeated measurements;

x_{ref} is the true or accepted reference value of concentration; and

s is the standard deviation of repeated measurements.

[EN 482]

3.5.4**precision**

closeness of agreement of results obtained by applying the same method several times under prescribed conditions

[ISO 6879]

NOTE Precision is often expressed in terms of the relative standard deviation.

3.5.5**true value**

value which characterizes a quantity perfectly defined in the conditions which exist when that quantity is considered

[ISO 3534-1]

NOTE The true value of a quantity is a theoretical concept and, in general, cannot be known exactly (see EN 1540).

4 Principle

4.1 A known volume of air is drawn through a filter to collect particulate lead and lead compounds. For personal sampling, a sampler designed to collect the inhalable fraction of airborne particles is used.

4.2 The filter and collected sample are subjected to a dissolution procedure in order to extract lead. The sample dissolution procedure may use one of three techniques: hot plate digestion, microwave digestion or ultrasonic extraction.

4.3 Sample solutions are analysed for lead content by aspirating into the oxidizing air-acetylene flame of an atomic absorption spectrometer equipped with a lead hollow-cathode lamp or electrodeless discharge lamp. Absorbance measurements are made at 283,3 nm, and analytical results are obtained by the analytical curve technique (see 6.1 of ISO 6955:1982). Potential interference by anions that form precipitates with lead is overcome by the addition of the disodium salt of ethylenediamine tetraacetic acid (EDTA) when necessary.

4.4 For accurate lead determination when the concentration of lead in the solution is low, the analysis may be repeated using electrothermal atomic absorption spectrometry. Aliquots of the test solution are injected into a

graphite furnace, and after drying and sample ashing stages, the sample is atomized electrothermally. Absorbance measurements are made at 283,3 nm with background correction, and results are obtained by the analytical curve technique (see 6.1 of ISO 6955:1982).

4.5 The results may be used for the assessment of workplace exposures to airborne particulate lead (see EN 689).

5 Reactions

In general, the overwhelming majority of particulate lead compounds that are commonly found in samples of workplace air are converted to water-soluble lead ions (Pb^{2+}) by the sample dissolution procedures described in 11.2. However, certain lead compounds, for example lead silicate, might not be dissolved. If necessary, a dissolution procedure employing hydrofluoric acid should be used to dissolve silicate lead. If there is any doubt about the effectiveness of these procedures for the dissolution of particulate lead compounds that may be present in the test atmosphere, then this shall be investigated before proceeding with the analytical method described in clause 11.

6 Requirement

The measuring procedure shall comply with any relevant international, european or national standard that specifies performance requirements for procedures for measuring chemical agents in workplace air (e.g. EN 482).

7 Reagents

During the analysis, use only reagents of recognized analytical grade, and only water as specified in 7.1.

7.1 Water, complying with the requirements for ISO 3696 grade 2 water (electrical conductivity less than 0,1 mS/m and resistivity greater than 0,01 $M\Omega \cdot m$ at 25 °C).

The concentration of lead in the water shall be less than 0,01 $\mu\text{g}/\text{ml}$.

It is recommended that the water used be obtained from a water purification system that delivers ultrapure water having a resistivity greater than 0,18 $M\Omega \cdot m$ (usually expressed by manufacturers of water purification systems as 18 $M\Omega \cdot \text{cm}$).

7.2 Nitric acid (HNO_3), concentrated, $\rho \approx 1,42 \text{ g}/\text{ml}$ (about 70 % mass fraction).

The concentration of lead shall be less than 0,01 $\mu\text{g}/\text{ml}$.

WARNING — Concentrated nitric acid is corrosive and oxidizing, and nitric acid fumes are irritant. Avoid exposure by contact with the skin or eyes, or by inhalation of fumes. Use suitable personal protective equipment (including suitable gloves, face shield or safety glasses, etc.) when working with the concentrated or dilute nitric acid, and carry out sample dissolution with concentrated nitric acid in open vessels in a fume hood.

7.3 Nitric acid, diluted 1 + 1.

Carefully add 500 ml of concentrated nitric acid (7.2) to 450 ml of water (7.1) in a 2-litre beaker. Swirl to mix, allow to cool and transfer to a 1-litre one-mark volumetric flask (8.6.1.4). Dilute to the mark with water, stopper and mix thoroughly.

7.4 Nitric acid, diluted 1 + 9.

Place approximately 800 ml of water (7.1) in a 1-litre one-mark volumetric flask (8.6.1.4). Carefully add 100 ml of concentrated nitric acid (7.2) to the flask and swirl to mix. Allow to cool, dilute to 1 litre with water and mix thoroughly.

7.5 Hydrofluoric acid (HF), concentrated, $\rho \approx 1,16 \text{ g}/\text{ml}$ (about 48 % mass fraction), if required, for digestion of samples containing lead silicates.

The concentration of lead in the HF shall be less than 0,1 µg/ml.

WARNING — Concentrated hydrofluoric acid and hydrogen fluoride vapour are extremely toxic and intensely corrosive, and diluted hydrofluoric acid can also cause serious and painful burns that might not be felt until up to 24 h after contact. Avoid exposure by contact with the skin or the eyes, or by inhalation of the vapour. Use of personal protection (e.g. impermeable gloves, face shield or safety glasses, etc.) is essential when working with concentrated or diluted hydrofluoric acid, and concentrated hydrofluoric acid should be used in a fume hood. It is essential that hydrofluoric acid antidote gel containing calcium gluconate is readily available to workers, both during and for 24 h after use of hydrofluoric acid.

7.6 Matrix modifier, $\text{NH}_4\text{H}_2\text{PO}_4$, $\text{Mg}(\text{NO}_3)_2$ or $\text{Pd}(\text{NO}_3)_2$, or a combination of these, if required, for analysis by electrothermal atomic absorption spectrometry.

7.7 Stock lead standard solution, 1 000 mg/l of lead.

Use a commercial standard solution with a certified lead concentration traceable to national standards. Observe the manufacturer's expiration date or recommended shelf life.

Alternatively, prepare a lead standard solution by one of the following procedures.

- Dissolve 1,598 g \pm 0,001 g of lead(II) nitrate [$\text{Pb}(\text{NO}_3)_2$], previously dried to constant mass at 110 °C and cooled in a desiccator, in 200 ml of 1 + 1 nitric acid (7.3). Quantitatively transfer the solution to a 1000 ml one-mark volumetric flask (8.6.1.4). Dilute to the mark with water (7.1), stopper and mix thoroughly. Store in a suitable container, e.g. a polypropylene bottle (8.6.2.2), for a maximum period of one year.
- Dissolve 1,000 g \pm 0,001 g of lead wire (99,9 % mass fraction Pb) in 200 ml of 1 + 1 nitric acid (7.3). Quantitatively transfer the solution into a 1 000 ml one-mark volumetric flask (8.6.1.4), dilute to the mark with water (7.1), stopper and mix thoroughly. Store in a suitable container, e.g. a polypropylene bottle (8.6.2.2), for a maximum period of one year.

7.8 Working lead standard solution, 1 mg/l of lead, if required, for analysis by electrothermal atomic absorption spectrometry.

Accurately pipette 100 µl of stock lead standard solution (7.7) into a 100 ml one-mark volumetric flask (8.6.1.4). Add 1 ml of concentrated nitric acid (7.2), dilute to the mark with water (7.1), stopper and mix thoroughly. Store in a suitable container, e.g. a polypropylene bottle (8.6.2.2), for a maximum period of one month.

7.9 Hydrogen peroxide (H_2O_2), approximately 30 % mass fraction solution, if required, for use in the hot-plate sample digestion method.

The concentration of lead in the hydrogen peroxide solution shall be less than 0,01 µg/ml.

7.10 Acetylene, if required, for use in analysis by flame atomic absorption spectrometry.

7.11 Air, compressed and filtered, if required, for use in analysis by flame atomic absorption spectrometry.

8 Apparatus

8.1 Inhalable samplers, designed to collect the inhalable fraction of airborne particles, complying with the provisions of EN 13205, for use when the exposure limits of interest apply to the inhalable fraction of airborne particles.

NOTE 1 In general, personal samplers for collection of the inhalable fraction of airborne particles do not exhibit the same size-selective characteristics if used for static (area) sampling.

NOTE 2 Some inhalable samplers are designed to collect the fraction of airborne particles on a filter, and any particulate matter deposited on the internal surfaces of the sampler is not of interest. Other inhalable samplers are designed such that airborne particles that pass through the entry orifice(s) match the inhalable convention, in which case particulate matter deposited on the internal surfaces of the sampler does form part of the sample. (Samplers of this second type generally incorporate an internal filter cassette or cartridge that can be removed from the sampler to enable this material to be easily recovered.) The operating

instructions supplied by the manufacturer should be consulted to find out whether particulate matter deposited on the internal surfaces of the sampler forms part of the sample.

8.2 Filters, of a diameter suitable for use with the samplers (see 8.1), with a collection efficiency of not less than 99,5 % for particles with a 0,3 μm diffusion diameter in accordance with 2.2 of ISO 7708:1995, with a minimum lead content (typically less than 0,1 $\mu\text{g Pb}$), and compatible with the selected sample preparation method.

NOTE See annex A for guidance on filter selection.

8.3 Sampling pumps with an adjustable flowrate and capable of maintaining the selected flowrate (between 1 l/min and 5 l/min for personal sampling pumps, and between 5 l/min and 400 l/min for high-volume sampling pumps) to within ± 5 % of the nominal value throughout the sampling period (see 10.1.2).

NOTE A flow-stabilized pump may be required to maintain the flowrate within the specified limits.

For personal sampling the pumps shall be capable of being worn by the worker without impeding normal work activity. Sampling pump flowmeters shall be calibrated using either a primary or secondary standard; if a secondary standard is used, it shall be calibrated using a primary standard.

The pump should have, as a minimum, the following features:

- an automatic control that keeps the volumetric flowrate constant in the case of a changing back-pressure;
- either a malfunction indicator which, following completion of sampling, indicates that the air flow has been reduced or interrupted during sampling; or an automatic cut-out, which stops the pump if the flowrate is reduced or interrupted;
- a facility for the adjustment of flowrate, such that it can only be actuated with the aid of a tool (e.g. screwdriver) or requires special knowledge for operation (e.g. via software), so as to preclude inadvertent readjustment of the flowrate during use.

An integral timer is a highly desirable additional feature.

EN 1232 and EN 12919 require that the performance of the pumps be such that:

- the pulsation of the flowrate does not exceed 10 %;
- a flowrate set within the nominal range does not deviate by more than ± 5 % from the initial value under increasing back-pressure;
- within the range of ambient temperatures from 5 °C to 40 °C, the flowrate measured under operating conditions does not deviate by more than ± 5 % from the flowrate at 20 °C;
- the operating time is at least 2 h, and preferably 8 h;
- the flowrate does not deviate by more than ± 5 % from the initial value during the operating time.

If the sampling pump is used outside the range of conditions specified in EN 1232 and/or EN 12919, appropriate action should be taken to ensure that the performance requirements are met. For instance, at sub-zero temperatures it might be necessary to keep the pump warm by placing it under the worker's clothes.

8.4 Flowmeter, portable, with an accuracy that is sufficient to enable the volumetric flowrate (see 10.1.1.2) to be measured to within ± 5 %.

The calibration of the flowmeter shall be checked against a primary standard, i.e. a flowmeter whose accuracy is traceable to national standards. If appropriate (see 10.1.3.1), record the atmospheric temperature and pressure at which the calibration of the flowmeter was checked.

It is recommended that the flowmeter used be capable of measuring the volumetric flowrate to within ± 2 % or better.

8.5 Ancillary equipment.

8.5.1 Flexible tubing, of a diameter suitable for making a leak-proof connection from the samplers to the sampling pumps.

8.5.2 Belts or harnesses, to which the sampling pumps can conveniently be fixed for personal sampling (except where the sampling pumps are small enough to fit inside worker's pockets).

8.5.3 Flat-tipped forceps, for loading and unloading filters into samplers.

8.5.4 Filter transport cassettes, or similar, if required to transport samples for laboratory analysis.

8.5.5 Barometer, suitable for measurement of atmospheric pressure, if required (see 10.1.3).

8.5.6 Thermometer, minimum temperature range of 0 °C to 50 °C, with graduated divisions of 1 °C or less, for measurement of atmospheric temperature.

For applications at temperatures below freezing, the range of the thermometer shall extend to the appropriate desired range.

8.6 Analytical or laboratory apparatus.

Ordinary laboratory apparatus, and:

8.6.1 Glassware, made of borosilicate glass 3.3 and complying with the requirements of ISO 3585.

It is preferable to reserve a set of glassware for analysis of lead by this method, in order to ensure that problems do not arise from incomplete removal of lead contamination by cleaning.

8.6.1.1 Beakers, of capacities between 50 ml and 150 ml, with watch-glasses to fit the beakers; for hot plate procedures.

8.6.1.2 One-mark pipettes, complying with the requirements of ISO 648.

8.6.1.3 Measuring cylinder, of capacity between 10 ml and 1 000 ml. (Also often referred to as a graduated cylinder.)

8.6.1.4 One-mark volumetric flasks, of capacities between 10 ml and 1 000 ml, complying with the requirements of ISO 1042.

8.6.2 Plastic labware.

8.6.2.1 Heatable beakers, beaker covers, etc., if required, made of a material that is resistant to corrosion by hydrofluoric acid, e.g. a fluorocarbon polymer such as polytetrafluoroethylene (PTFE), and suitable for performing dissolutions using hydrofluoric acid.

8.6.2.2 Polypropylene bottles, of capacities from 100 ml to 1 000 ml.

8.6.3 Piston-operated volumetric instruments, complying with the requirements of ISO 8655-1 and tested in accordance with ISO 8655-6; **pipetters**, complying with the requirements of ISO 8655-2, as an alternative to one-mark pipettes, for the preparation of standard solutions, calibration solutions and dilution of samples; and **dispensers**, complying with the requirements of ISO 8655-5, for dispensing acids.

8.6.4 Hot plate, thermostatically controlled, capable of maintaining a surface temperature of approximately 150 °C, for hot-plate procedures.

The efficiency of thermostating of hot plates is sometimes deficient, and the surface temperature can also vary considerably with position on hot plate with large surface areas. It is therefore recommended that the performance of the hot plate be characterized prior to use.

8.6.5 Microwave digestion apparatus.

8.6.5.1 General

Ensure that manufacturer's safety recommendations are followed.

NOTE 1 The specified method is for closed vessel microwave digestion systems with a temperature control system. Microwave digestion systems that are equipped only with a pressure control system and/or with lower pressure vessels may be used provided that a suitable sample dissolution procedure is developed and a prior assessment of dissolution efficiency is carried out.

NOTE 2 Open-vessel microwave digestion systems can give results equivalent to closed-vessel microwave digestion systems. They may therefore be used provided that a suitable sample dissolution procedure is developed and a prior assessment of dissolution efficiency is carried out.

8.6.5.2 Microwave digestion system, designed for closed-vessel sample digestion in the laboratory, with power output regulation, fitted with a temperature control system capable of sensing the temperature to within ± 2 °C and automatically adjusting the microwave power output within 2 s.

The microwave cavity shall be corrosion-resistant and well ventilated, with all electronics protected against corrosion to ensure safe operation.

CAUTION — Do not use domestic (kitchen) microwave ovens, since there are very significant hazards associated with their use for the procedure described in this International Standard. Acid vapours released into the cavity can corrode safety devices that prevent the magnetron from shutting off when the door is opened, potentially exposing the operator to microwave energy. Also, the fumes generated can be extremely hazardous.

NOTE A pressure control system is also very useful, since it provides a safeguard against the possibility of sample loss due to excessive pressure build-up and partial venting of the sample vessels.

8.6.5.3 Vessels, designed for carrying out microwave digestions, capable of withstanding a temperature of 180 °C, and with an internal volume of at least 50 ml.

The vessels shall be transparent to microwave energy, and shall be capable of withstanding internal pressures up to at least 3 000 kPa (435 psi) or greater, and temperatures up to at least 180 °C or greater. Closed vessels shall also be equipped with a safety relief valve or disc that will prevent vessel rupture or ejection of the vessel cap. Such vessels consist of an inner liner and cover made of a microwave-transparent and chemically resistant material [usually a fluorocarbon polymer such as tetrafluoromethoxyl polymer (TFM)], which contains and isolates the sample solution from a high-strength, outer pressure vessel structure. Other types of sample vessel designed to operate at equivalent or higher temperatures or pressures, or both, may be used.

CAUTION — For closed-vessel designs, the material from which the outer vessels are made is usually not as chemically resistant as the liner material. Since the outer vessels provide the strength required to withstand the high pressures within the inner liners, they shall be inspected regularly to check for any chemical or physical degradation.

8.6.6 Ultrasonic bath (sonicator), for performing ultrasonic extractions, capable of delivering sufficient power to effect the quantitative dissolution of particulate lead under the conditions described in 11.2.5 (typically 1 W/cm² power density or greater).

8.6.7 Plastic centrifuge tubes, 50 ml, with screw caps (for ultrasonic procedure).

8.6.8 Atomic absorption spectrometer, fitted with an air-acetylene burner supplied with compressed air and acetylene, and equipped with either a lead hollow cathode lamp or electrodeless discharge lamp [4, 5]. If sample dissolution is carried out with the aid of hydrofluoric acid (see 11.2.3.3 and 11.2.4.2), the atomic absorption spectrometer shall be hydrofluoric acid-compatible. If electrothermal atomic absorption is to be carried out, the atomic absorption spectrometer shall be capable of carrying out simultaneous background correction at 283,3 nm, either by using a continuum source such as a deuterium lamp to measure non-specific attenuation (see for example 5.1.5 of ISO 6955:1982), or by using Zeeman or Smith-Hieftje background correction systems [6].

8.6.9 Electrothermal atomiser, fitted with a solid, pyrolytic graphite platform mounted in a pyrolytically-coated graphite tube, supplied with argon purge gas, and equipped with an autosampler capable of injecting microlitre volumes onto the platform.

NOTE Some manufacturers of atomic absorption spectrometers use an alternative design of electrothermal atomiser to achieve a constant temperature environment during atomisation, and some use aerosol deposition as a means of sample introduction. The use of such accessories is acceptable, provided satisfactory method performance is verified. Likewise, atomisers made from heat-resistant metal, e.g. tungsten, might also be suitable.

8.6.10 Analytical balance, capable of weighing to $\pm 0,1$ mg, if required, for use in preparation of stock standard lead solution.

8.6.11 Disposable gloves, for prevention of sample contamination.

8.6.12 Forceps, flat-tipped, for loading and unloading of filters into and out of samplers.

9 Occupational exposure assessment

9.1 Assessment strategy

Refer to relevant International or national Standards (e.g. EN 689, ASTM E1370) for guidance on how to develop an appropriate assessment strategy.

9.2 Measurement strategy

9.2.1 General

Refer to relevant International or national Standards (e.g. EN 689, ASTM E1370) for general guidance on measurement strategy.

9.2.2 Personal sampling

Exposure of workers to lead shall normally be determined by personal sampling, since the concentration of lead and lead compounds in the breathing zone is usually higher than their background levels in the workplace.

9.2.3 Static (area) sampling

Static (area) sampling may be carried out, if appropriate, to assess the exposure of workers in a situation where personal sampling is not possible; to characterise the background level of lead in the workplace to give an indication of the efficiency of ventilation or other engineering controls; or to provide information on the location and intensity of an emission source.

9.3 Selection of measurement conditions and measurement pattern

9.3.1 General

9.3.1.1 The sampling procedure shall be devised to cause the least possible interference with the worker and the normal performance of the job, and to provide samples that are representative of normal working conditions and that are compatible with the analytical method.

9.3.1.2 The pattern of sampling shall take into consideration practical issues, such as the nature of the measurement task and the frequency and duration of particular work activities.

9.3.2 Screening measurements of variation of concentration in time and/or space

Screening measurements of variation of concentration in time and/or space may be carried out in the initial stages of a survey to identify locations and periods of elevated exposure, and to set the duration and frequency of sampling for measurements for comparison with limit values.

NOTE For making screening measurements of variation of concentration in time and/or space, the sampling time used is normally between 5 min and 30 min.

9.3.3 Screening measurements of time-weighted average concentration and worst-case measurements

Screening measurements of time-weighted average concentration may be carried out in the initial stages of a survey to assess the effectiveness of control measures. This may involve sampling during representative work episodes to obtain clear information about the level and pattern of exposure, or worst-case measurements can be made.

9.3.4 Measurements for comparison with limit values and periodic measurements

For making long-term measurements, samples shall be collected for the entire working period or during a number of representative work episodes [3].

NOTE The best estimate of long-term exposure is obtained by taking samples for the entire working period, but this is often not practicable or not desirable (e.g. because of the possibility of overloading the filter).

10 Sampling

10.1 Preliminary considerations

10.1.1 Selection and use of samplers

10.1.1.1 Select samplers (8.1) designed to collect the inhalable fraction of airborne particles, as defined in ISO 7708.

If possible, the samplers selected should be manufactured from conducting material, since samplers comprised of non-conducting material have electrostatic properties that can influence representative sampling.

10.1.1.2 Use the samplers at their design flowrate and in accordance with the manufacturer's instructions, so that they collect the inhalable fraction of airborne particles.

10.1.2 Sampling period

10.1.2.1 Select a sampling period long enough to ensure that the amount of lead collected is adequate to enable lead-in-air concentrations to be determined at the required level (see 9.3).

10.1.2.2 In calculating the minimum sampling time required it is necessary to consider the selected flowrate and the lower limit of the analytical working range of the method [7].

10.1.2.3 When high concentrations of airborne particles are anticipated, select a sampling period that is not so long as to risk overloading the filter with particulate matter.

NOTE If filter overloading is an observed or suspected problem and it is desired to sample for the entire working day, it might be necessary to collect consecutive samples [8].

10.1.3 Temperature and pressure effects

10.1.3.1 Expression of results

Consider whether it is necessary to recalculate the concentration of lead in air to reference conditions (such as in high altitude situations). If so, measure and record the atmospheric temperature and pressure at the start and at the end of the sampling period (see 10.4.1 and 10.4.2) and use the equation given in clause B.2 to apply the necessary correction.

NOTE The concentration of lead in air is generally stated for actual environmental conditions (temperature, pressure) at the workplace during the sampling period.

10.1.3.2 Effect of temperature and pressure on flowrate measurements

Refer to the manufacturer's instructions to determine if the indicated volumetric flowrate of the flowmeter (8.4) is dependent upon temperature and pressure. Consider whether the difference between the atmospheric temperature and pressure at the time of calibration of the flowmeter and during sampling is likely to be great enough to justify making a correction to take this into account, e.g. if the error could be greater than $\pm 5\%$. If a correction is necessary, measure and record the atmospheric temperature and pressure at which the calibration of the flowmeter was checked (see 8.4) and measure and record the atmospheric temperature and pressure at the start and at the end of the sampling period (see 10.4.1 and 10.4.2).

NOTE An example of temperature and pressure correction for the indicated volumetric flowrate is given in clause B.1 for a constant pressure drop, variable area, flowmeter.

10.2 Preparation of sampling equipment

10.2.1 Cleaning of samplers

Unless disposable filter cassettes are used, clean the samplers (8.1) before use. Disassemble the samplers, soak in detergent solution, rinse thoroughly with water, wipe with absorbent tissue and allow to dry before reassembly. Alternatively, use a laboratory washing machine.

10.2.2 Loading the samplers with filters

Load clean samplers (see 10.2.1) with filters (8.2), label each sampler so that it can be uniquely identified and seal with its protective cover or plug to prevent contamination.

Alternatively, commercially available pre-loaded filter cassettes may be used.

10.2.3 Setting the volumetric flowrate

Perform the following in a clean area, where the concentration of lead is low:

Connect each loaded sampler (see 10.2.2) to a sampling pump (8.3) using flexible tubing (8.5.1), ensuring that no leaks can occur. Remove the protective cover or plug from each sampler, switch on the sampling pump, attach the flowmeter (8.4) to the sampler so that it measures the flow through the sampler inlet orifice(s), and set the required volumetric flowrate (see 10.1.1.2). Switch off the sampling pump and seal the sampler with its protective cover or plug to prevent contamination during transport to the sampling position.

If necessary, allow the sampling pump operating conditions to stabilize before setting the volumetric flowrate.

10.2.4 Blanks

Retain as blanks one unused loaded sampler from each batch of ten prepared, subject to a minimum of three. Treat these in the same manner as those used for sampling in respect of storage and transport to and from the sampling position, but draw no air through the filters.

10.3 Sampling position

10.3.1 Personal sampling

Position the sampler in the worker's breathing zone, as close to the mouth and nose as is reasonably practicable, e.g. fastened to the worker's lapel. Attach the sampling pump to the worker in a manner that causes minimum inconvenience, e.g. to a belt (8.5.2) around the waist, or place it in a convenient pocket.

10.3.2 Static (area) sampling

10.3.2.1 If static sampling is carried out to assess the exposure of a worker in a situation where personal sampling is not possible (e.g. due to the need to sample at a volumetric flowrate higher than the design flowrate of available personal samplers), position the sampler in the immediate vicinity of the worker and at breathing height. If in doubt, take the sampling position to be the point where the risk of exposure is considered to be greatest.

10.3.2.2 If static sampling is carried out to characterize the background level of lead in the workplace, select a sampling position that is sufficiently remote from the work processes, such that results will not be directly affected by lead from emission sources.

10.4 Collection of samples

10.4.1 When ready to begin sampling, remove the protective cover or plug from the sampler and switch on the sampling pump. Record the time and volumetric flowrate at the start of the sampling period. If the sampling pump is fitted with an integral timer, check that this is reset to zero. If appropriate (see 10.1.1.2), measure the atmospheric temperature and pressure at the start of the sampling period using the thermometer (8.5.6) and barometer (8.5.5), and record the measured values

NOTE If the temperature or pressure at the sampling position is different from that where the volumetric flowrate was set (see 10.2.3), the volumetric flowrate could change and it might need to be re-adjusted before sampling.

10.4.2 At the end of the sampling period (see 10.1.2), record the time and calculate the duration of the sampling period. Check the malfunction indicator and/or the reading on the integral timer, if fitted, and consider the sample to be invalid if there is evidence that the sampling pump was not operating properly throughout the sampling period. Measure the volumetric flowrate at the end of the sampling period using the flowmeter (8.4), and record the measured value. If appropriate (see 10.1.3), measure the atmospheric temperature and pressure at the end of the sampling period using the thermometer (8.5.6) and barometer (8.5.5), and record the measured values.

10.4.3 Carefully record the sample identity and all relevant sampling data (see clause 14). Calculate the mean volumetric flowrate by averaging the volumetric flowrates at the start and at the end of the sampling period and, if appropriate (see 10.1.3), calculate the mean atmospheric temperature and pressure. Calculate the volume of air sampled, in litres, at atmospheric temperature and pressure, by multiplying the mean flowrate in litres per minute by the duration of the sampling period in minutes.

10.5 Transportation

10.5.1 For samplers which collect airborne particles on the filter (see NOTE 2 in 8.1), remove the filter from each sampler, place in a labelled filter transport cassette (8.5.4) and close with a lid. Take particular care to prevent the collected sample from becoming dislodged from heavily loaded filters. Alternatively, transport samples to the laboratory in the samplers in which they were collected.

10.5.2 For samplers which have an internal filter cassette (see NOTE 2 in 8.1), remove the filter cassette from each sampler and fasten with its lid or transport clip.

10.5.3 For samplers of the disposable cassette type, transport samples to the laboratory in the samplers in which they were collected.

10.5.4 Transport the samples (10.5.1 to 10.5.3) to the laboratory in a container that has been designed to prevent damage to the samples in transit and which has been labelled to assure proper handling.

10.5.5 Follow sampling chain of custody procedures to ensure sample traceability. Ensure that the documentation which accompanies the samples is suitable for a "chain of custody" to be established (see, for example, ASTM D4840-88).

11 Analysis

11.1 Cleaning of glassware and plasticware

11.1.1 Perform all of the following while wearing gloves.

11.1.2 Before use, clean all glassware, microwave digestion vessels, and plasticware to remove any residual grease or chemicals by first soaking in laboratory detergent solution and then rinsing thoroughly with water (7.1).

11.1.3 After initial cleaning with detergent and water, clean all beakers with nitric acid. This can be accomplished either by soaking for a minimum of 24 h in concentrated nitric acid (7.2), or by the following procedure. Fill beakers to one-third capacity with concentrated nitric acid (7.2), and then heat them on a hot plate with a surface temperature of 140 °C in a fume hood until most of the liquid has evaporated, and allow to cool. Rinse beakers thoroughly with water (7.1).

11.1.4 Glassware that has been previously subjected to the entire cleaning procedure described in the previous steps, and which has been reserved for the analysis of lead, can be cleaned adequately by rinsing with 1 + 9 nitric acid (7.4) and then with water (7.1).

11.1.5 Before use, clean polypropylene bottles, microwave digestion vessels and other plasticware by soaking them in 1 + 9 nitric acid (7.4) for at least 24 h and then rinse thoroughly with water (7.1).

NOTE Plasticware (possibly disposable) can be received in clean condition directly from the vendor, thereby precluding the need for cleaning prior to use.

11.2 Preparation of sample and blank solutions

11.2.1 General

Perform all of the following preparations while wearing gloves.

11.2.2 Selection of sample dissolution method

Prepare samples and blanks for analysis using one of the three sample preparation methods described below: either hot-plate digestion, microwave digestion or ultrasonic extraction.

11.2.3 Hot-plate digestion method

11.2.3.1 Open the samplers, sampler filter cassettes or transport filter cassettes (see 10.5), and transfer each filter sample or blank into a clean, labelled 50 ml beaker (8.6.1.1) using flat-tipped forceps (8.5.3). If the sampler used was of a type in which airborne particles deposited on the internal surfaces of the sampler form part of the sample, wash

any particulate matter adhering to the internal surfaces into the beaker using a minimum volume of 1 + 9 nitric acid (7.4).

11.2.3.2 To each beaker, add 3 ml of concentrated nitric acid (7.2) and 1 ml of hydrogen peroxide (7.9), and cover with a watch-glass.

11.2.3.3 Heat on a hot plate (8.6.4) with a surface temperature of approximately 140 °C in a fume hood, and allow the solution to evaporate until the final solution volume is reduced to approximately 1 ml. Avoid taking to dryness. Remove beakers from the hot plate and allow to cool.

NOTE The exact hot-plate temperature is not critical. A temperature of 140 °C is used because it is high enough to enable the liquid to be evaporated at an acceptable rate. This temperature is also useful for minimizing the risk of taking samples to dryness.

The use of hydrofluoric acid (HF) (7.5) is required to dissolve silicate lead. If the material in the test atmosphere is believed to contain a significant amount of silicate material, its dissolution can be facilitated by adding 1 ml of HF at the same time as the nitric acid. However, it will be necessary to use heatable beakers and beaker covers, etc. that are made of plastic that is resistant to corrosion by HF, e.g. a fluorocarbon polymer such as PTFE.

11.2.3.4 Carefully rinse each watch-glass and the sides of each beaker with water, and transfer each solution quantitatively to a 10 ml one-mark volumetric flask (8.6.1.4). If necessary, remove any undissolved particulate by filtration or centrifugation. Dilute to the mark of the volumetric flask with water (7.1), seal the flask with a stopper, and mix thoroughly.

11.2.4 Microwave digestion method

11.2.4.1 Open the samplers, sampler filter cassettes or transport filter cassettes (see 10.5), and transfer each filter into the clean liner of a labelled microwave digestion vessel (8.6.5.2) using flat-tipped forceps (8.5.3). Follow the same procedure for blank filters. If the sampler used was of a type in which airborne particles deposited on the internal surfaces of the sampler form part of the sample, wash any particulate matter adhering to the internal surfaces into the vessel liner using a minimum volume of water (7.1).

11.2.4.2 Carefully add 5 ml of concentrated nitric acid (7.2) to the inside of liner of the microwave digestion vessel containing the filter sample or blank. Seal the vessels.

The use of hydrofluoric acid (7.5) is required to dissolve lead silicates. If the material present in the test atmosphere is believed to contain a significant amount of silicate material, its dissolution can be facilitated by adding 1 ml of hydrofluoric acid at the same time as the nitric acid.

11.2.4.3 Load the vessels into the microwave oven (8.6.5.1) according to manufacturer's instructions. Vessels containing samples shall be evenly and symmetrically placed in the microwave oven.

Even, symmetrical spacing of vessels is needed to ensure uniform microwave heating of all vessel solutions.

11.2.4.4 Program the microwave digestion system to reach 180 °C in less than 10 min, and then hold at this temperature for 15 min.

NOTE If hydrofluoric acid is used to dissolve the samples and the temperature sensor is not resistant to attack by this acid, the vessel in which the temperature sensor is fitted should contain a filter blank in which an equal volume of nitric acid is substituted for the hydrofluoric acid used for dissolution of the samples.

11.2.4.5 At the end of the digestion period, remove the vessels from the microwave oven, place them in a fume hood, and allow the solutions to cool to room temperature.

11.2.4.6 For closed vessels, carefully detach the vent tubing, and carefully shake the vessels to vent any excess gas pressure that may be present inside the vessels. Carefully open each sample vessel.

11.2.4.7 Quantitatively transfer the contents of each vessel to 10 ml-one-mark volumetric flasks (8.6.1.4). Carefully rinse each vessel with water, and dilute to volume with water (7.1). If necessary, remove any undissolved particles by filtration or centrifugation. Seal each flask with a stopper and mix thoroughly.

11.2.5 Ultrasonic extraction method

11.2.5.1 The following method is not applicable to samples containing silicates. In such a case, use the method employing hydrofluoric acid described in annex D of ISO 15202-2:2001.

11.2.5.2 Open the samplers, sampler filter cassettes or transport filter cassettes (see 10.5), and transfer each filter sample or blank into a clean 50 ml centrifuge tube (8.6.7) using flat-tipped forceps (8.5.3). Label each centrifuge tube with a unique identifier. If the sampler used was of a type in which airborne particles deposited on the internal surfaces of the sampler form part of the sample, wash any particulate matter adhering to the internal surfaces into the centrifuge tube using a minimum volume of water (7.1). Using a clean glass or plastic rod, push the filter to the bottom of the centrifuge tube.

11.2.5.3 Introduce 10 ml of 1 + 9 nitric acid (7.4) into each centrifuge tube containing a filter sample or blank, and cap each tube.

11.2.5.4 Place each centrifuge tube upright in an ultrasonic bath (8.6.6), and ensure that the water level within the bath is at or above the level of liquid within the tube.

NOTE Depending on the size of the ultrasonic bath, many centrifuge tubes may be immersed in the bath at one time. A custom rack for the centrifuge tubes can be purchased or constructed to allow for the regular and orderly placement of multiple tubes in the sonicator bath.

11.2.5.5 Apply ultrasonic energy to the acid-immersed filter samples for a minimum of 30 min.

11.2.5.6 Remove centrifuge tubes from the bath. Keep tubes in upright position, and allow to cool to room temperature.

11.3 Instrumental analysis

11.3.1 Selection of analytical line

The 283,3 nm lead analytical line shall be used for making absorbance measurements.

NOTE The most sensitive lead line is at 217,0 nm. However, this line is subject to possible spectral interference from antimony, and the significant spectral background at 217,0 nm makes correction for non-specific attenuation (see 5.1.5 of ISO 6955:1982) essential at this wavelength. The 283,3 nm line exhibits somewhat lower sensitivity than the 217,0 nm line, but it is not subject to spectral interference. In addition, while the detection limits obtained are dependent upon the instrument used, absorbance measurements made at 283,3 nm generally have a better signal-to-noise ratio than those made at 217,0 nm, and hence a better detection limit.

11.3.2 Flame atomic absorption spectrometry

11.3.2.1 Instrument set-up

Set up the atomic absorption spectrometer (8.6.8) to make absorbance measurements at 283,3 nm, following the manufacturer's instructions for specific instrument operating parameters. Use a lead hollow cathode lamp or electrodeless discharge lamp and an oxidising air-acetylene flame. Allow an appropriate warm-up period for the source lamp.

11.3.2.2 Preparation of calibration solutions

11.3.2.2.1 Use 1 + 9 nitric acid (7.4) as the solvent blank (see 5.4.2 of ISO 6955:1982).

11.3.2.2.2 Prepare at least four calibration solutions, including a blank calibration solution, to cover a suitable concentration range, e.g. from 0 µg/ml to 20 µg/ml of lead. Accurately pipette appropriate volumes of stock lead standard solution (7.7) into separate, labelled 100 ml one-mark volumetric flasks (8.6.1.4). Dilute to the mark with 1 + 1 nitric acid (7.3) for test solutions prepared by the microwave digestion method, or with 1 + 9 nitric acid (7.4) for

test solutions prepared by the hot plate or ultrasonic sample digestion methods. Stopper and mix thoroughly. Prepare these calibration solutions fresh daily.

NOTE The concentration range of calibration solutions is given as a guide. The upper limit of the working range is dependent upon which wavelength is used, and it is also governed by instrumental factors that affect sensitivity and the linearity of the calibration. Accordingly, the range of the set of calibration solutions may be varied, but when making any changes, ensure that the response of the spectrometer over the alternative range of concentrations selected is such that it is linear.

11.3.2.3 Calibration measurements

11.3.2.3.1 Adjust the spectrometer zero while aspirating the solvent blank (see 11.3.2.2.1) into the flame. Then aspirate the calibration solutions into the flame, and make absorbance measurements for each solution.

Use of an autosampler is recommended, since precision is maximized and the volume of solution consumed is minimized.

11.3.2.3.2 Analyse all blank solutions, and calculate the mean concentration of the blank solutions.

11.3.2.4 Calibration function

For instruments controlled by a microprocessor or personal computer, use a suitable algorithm to generate the calibration function. For instruments without this capability, prepare a calibration graph by plotting the absorbance of the calibration solutions versus the concentration of lead ($\mu\text{g}/\text{ml}$) in the respective solutions.

In general, it is best to work within the linear range of the calibration, where absorbance is proportional to the lead concentration in solution.

11.3.2.5 Determination

11.3.2.5.1 Adjust the spectrometer zero while aspirating the solvent blank (see 11.3.2.2.1) into the flame. (Repeat this procedure regularly throughout the determination and readjust the zero if the baseline drifts.) Then aspirate the sample and blank test solutions into the flame and make absorbance measurements for each solution. For instruments controlled by a microprocessor or personal computer, use the calibration function to calculate the concentration of lead in the test solutions, and obtain a direct read-out of the results in concentration units. For instruments without this capability, determine the concentration of lead in the test solutions from the calibration graph.

11.3.2.5.2 Aspirate a mid-range calibration solution after each five to ten test solutions and make an absorbance measurement. If this indicates that the sensitivity has changed by more than $\pm 5\%$, take one of the following corrective measures: either use the available software facilities of the microprocessor or personal computer to correct for the sensitivity change ("reslope" facility); or suspend analysis and recalibrate the spectrometer. In either case, reanalyse the test solutions that were analysed during the period in which the sensitivity change occurred.

11.3.2.5.3 If high concentrations of lead are found, dilute the sample test solutions to bring the concentration within the calibration range. Make all dilutions so that the final nitric acid concentration is 1 + 9, and record the dilution factor DF .

11.3.2.5.4 Calculate the mean lead concentration of the blank test solution.

11.3.2.5.5 If the concentration of lead in the sample test solutions is less than 0,5 $\mu\text{g}/\text{ml}$, consider repeating the analysis using electrothermal atomic absorption spectrometry, since this technique gives more precise measurements at low concentrations.

11.3.3 Electrothermal atomic absorption spectrometry

11.3.3.1 General

Lead is present ubiquitously in the environment, and therefore it is imperative that strict standards of cleanliness are observed to avoid contamination of labware. This is particularly important when carrying out electrothermal atomic absorption spectrometry since the technique exhibits a very low detection limit. Ensure that all glassware is cleaned thoroughly before use, and autosampler cups are stored in 1 + 9 nitric acid until required.

11.3.3.2 Preparation of working calibration solutions

11.3.3.2.1 Prepare a working calibration solution at a concentration of 2,5 ng/ml of lead. Accurately pipette 250 μ l of working lead standard solution (7.8) into a 100 ml one-mark volumetric flask (8.6.1.4). Dilute to the mark with 1 + 1 nitric acid (7.3) for test solutions prepared by the microwave digestion method, or with 1 + 9 nitric acid (7.4) for test solutions prepared by the hot-plate or ultrasonic sample digestion methods. Stopper and mix thoroughly. Prepare this solution fresh weekly.

11.3.3.2.2 Prepare a working calibration blank solution following the procedure in the preceding paragraph, but omitting the 250 μ l of working lead standard solution. Prepare this solution fresh weekly.

11.3.3.2.3 For instruments controlled by a microprocessor or personal computer, use a suitable algorithm to generate the calibration function. For instruments without this capability, prepare a calibration graph by plotting the absorbance of the calibration solutions versus the concentration of lead, in micrograms per millilitre, in the respective solutions.

In general it is best to work in the linear range of the calibration, where absorbance is proportional to the concentration of lead in solution.

11.3.3.3 Calibration and determination

11.3.3.3.1 Set up the atomic absorption spectrometer (8.6.8) and the electrothermal atomiser (8.6.9) to measure lead at a wavelength of 283,3 nm, using background correction to correct for non-specific attenuation (see for example 5.1.5 of ISO 6955:1982). Follow the manufacturer's instructions for specific operating parameters. Allow a suitable warm-up period for the hollow cathode lamp or equivalent source.

NOTE The operating parameters for electrothermal atomic absorption vary considerably between different instruments, much more so than for flame atomic absorption spectrometry.

11.3.3.3.2 Program the autosampler to prepare matrix-modified calibration solutions, sample test solutions and blank test solutions *in situ* on a pyrolytic graphite platform mounted in the pyrolytically coated graphite tube of the electrothermal atomizer. Prepare at least four matrix-modified calibration solutions to cover the range 0 ng/ml to 50 ng/ml of lead using the working calibration solution (11.3.3.2.1), the working calibration blank solution (11.3.3.2.2), and matrix modifier (7.6). Also prepare matrix-modified sample and blank test solutions using the unmodified sample and blank solutions and matrix-modifier solution. Matrix-modified calibration and test solutions shall be prepared by an autosampler or manually by means of one-mark volumetric flasks.

NOTE The procedure described above may be varied to accommodate the use of electrothermal atomizers of alternative design.

11.3.3.3.3 Set up the analytical sequence in the microprocessor or personal computer interfaced to the electrothermal atomic absorption spectrometric instrument. Specify an appropriate number of replicate analyses for each solution, and insert a calibration blank solution and a mid-range calibration solution after each five to ten test solutions in order to monitor for baseline drift and sensitivity change, respectively.

11.3.3.3.4 Place the working calibration solution (11.3.3.2.1), the working calibration blank solution (11.3.3.2.2), the matrix modifier (7.6), and the unmodified sample and blank solutions in separate acid-washed autosampler cups and position as appropriate in the autosampler carousel. Analyse the matrix-modified calibration and test solutions, using the microprocessor or personal computer software to generate a calibration. Obtain a direct readout of sample and blank results in nanograms of lead per millilitre.

11.3.3.5 If significant baseline drift is observed during the course of analysis, or if the sensitivity changes by more than $\pm 5\%$, take one or more of the following corrective measures: either use the available software facilities of the microprocessor or personal computer to correct for sensitivity change ("reslope" facility); or suspend analysis and recalibrate the spectrometer. In either case, reanalyse the solutions that were analysed during the period in which the sensitivity change occurred.

11.3.3.6 If concentrations of lead above the upper limit of the linear calibration range are found, dilute the sample test solutions in order to bring them within the range of the calibration, and repeat the analysis. Make all dilutions so that the final nitric acid concentration is 1 + 1 or 1 + 9, as appropriate, and record the dilution factor (*DF*). Alternatively, analyse a smaller aliquot of sample, and make a correction for the amount of sample that is analysed.

11.3.3.7 Calculate the mean concentration of lead in the blank test solutions.

11.4 Estimation of the instrumental detection limit

11.4.1 Estimate the instrumental detection limit under the working analytical conditions following the procedure described below, and repeat this exercise whenever the experimental conditions are changed.

11.4.2 Prepare test solutions at a concentration of 0,1 $\mu\text{g}/\text{ml}$ of lead for flame atomic absorption analysis or 1 ng/ml for electrothermal atomic absorption analysis by diluting the working lead standard solution (7.8). Make these dilutions so that the final nitric acid concentration is 1 + 9.

11.4.3 Make at least 20 absorbance measurements on the test solution and calculate the instrumental detection limit as three times the sample standard deviation of the mean concentration value (see for example 6.2.3 of ISO 6955:1982).

NOTE The limit of detection calculated from results using this procedure is an instrumental detection limit. This is of use in identifying changes in instrument performance, but it is not a method detection limit [7]. The instrumental detection limit is likely to be unrealistically low because it only takes into account the variability between individual instrumental readings; determinations made on one solution do not take into consideration contributions to variability from the matrix or sample.

11.5 Estimation of the method detection limit

11.5.1 Estimate the method detection limit under the working analytical conditions following the procedure described in 11.5.2 through 11.5.3, and repeat this exercise whenever the experimental conditions are changed significantly.

11.5.2 Fortify at least ten filters (8.2) with lead near the anticipated detection limit, e.g. 1 μg of lead for flame atomic absorption analysis or 0,01 μg of lead for electrothermal atomic absorption analysis, by spiking the filter with 0,1 ml of a suitable calibration solution (7.8) diluted by an appropriate factor with 1 + 9 nitric acid (7.4).

11.5.3 Make atomic absorption measurements on the test solutions derived from each spiked filter (11.5.2) (after carrying out digestion of the filters), and calculate the method detection limit as three times the sample standard deviation of the mean concentration value.

NOTE An alternative procedure for estimating the method detection limit involves the analysis of filter samples fortified with the analyte of interest at values spanning the predicted detection limit [7].

11.6 Quality control

11.6.1 General

Quality control (QC) samples to process with each batch of field samples are summarized below.

11.6.2 Reagent blanks and media blanks

Carry reagent blanks (water and reagents) and media blanks (unspiked filters) throughout the entire sample preparation and analytical process to determine whether the samples are being contaminated from laboratory activities. Process reagent blanks according to a frequency of at least one per 20 samples or a minimum of one per batch.

11.6.3 Spiked samples and spiked duplicate samples

11.6.3.1 Process these samples on a routine basis to estimate the method accuracy on the sample batch, expressed as a percent recovery relative to the true spiked value. Spiked samples and spiked duplicate samples consist of filters to which known amounts of analyte were added. (This can be accomplished by spiking known volumes of known concentrations of lead solution at amounts within the dynamic range of the instrument. The lead solution used shall be prepared from a stock standard solution from a different source than that used for preparing the calibration solutions.) Process these QC samples at a frequency of at least 1 per 20 samples or minimum of one per batch.

11.6.3.2 Monitor the performance of the method by plotting control charts of the relative percent recoveries and of the relative percent differences between the spiked samples and the spiked duplicate samples. If QC results indicate that the method is out of control, investigate the reasons for this, take corrective action and reanalyse the samples if necessary. See [9] for general guidance on the use of quality control charts.

11.6.4 Certified reference materials

Certified reference materials (CRMs) for lead shall be analysed prior to routine use of the method, and periodically thereafter, to establish that the percent recovery relative to the certified value is satisfactory. Suitable CRMs are available from many sources. A minimum of one CRM sample shall be analysed at least six times quarterly.

11.6.5 External quality assessment

If laboratories carry out lead in air analysis on a regular basis, it is strongly recommended that they participate in a relevant external quality assessment scheme or proficiency testing scheme.

12 Expression of results

12.1 Calculation

Calculate the mass concentration of lead in the air sample, ρ [Pb], in milligrams per cubic metre, at ambient conditions, using the equation:

$$\rho \text{ [Pb]} = \frac{(\rho \text{ [Pb]}_1 \cdot V_1 \cdot DF) - (\rho \text{ [Pb]}_0 \cdot V_0)}{V}$$

where

$\rho \text{ [Pb]}_0$ is the mean lead concentration, in micrograms per millilitre, in the blank test solutions;

$\rho \text{ [Pb]}_1$ is the lead concentration, in micrograms per millilitre, in the sample test solution;

V is the volume, in litres, of the air sample;

V_0 is the volume, in millilitres, of the blank test solution;

V_1 is the volume, in millilitres, of the sample test solution; and

DF is the dilution factor ($DF = 1$ in the absence of dilution).

12.2 Method performance

12.2.1 Sample collection

A collection efficiency of 1,00 was determined for the filter collection step for laboratory-generated lead nitrate aerosols and for lead fume [10].

12.2.2 Hot-plate digestion and flame atomic absorption spectrometry

12.2.2.1 The detection limit of flame atomic absorption measurements depends in part on the instrument used. However, the detection limit of the method has been estimated to be 0,25 μg per sample, and the precision of the measurement procedure was $< 0,1$ for samples in the range 0,9 μg to 2,25 μg and $< 0,03$ for samples in the range 3,6 μg to 288 μg [4]. No bias has been identified. The applicable range is 1 μg to 200 μg Pb per sample, without dilution.

12.2.2.2 In tests using laboratory-generated lead nitrate aerosols [10], the coefficient of variation for a similar procedure was found to be 0,072 for the overall sampling and analytical method in the range 130 μg to 400 μg Pb/m³. The bias of the method was found to be insignificant. Also, data from interlaboratory proficiency testing for samples of paint, soil and dust [11] and [12] have indicated insignificant bias.

NOTE If the sample dissolution procedure using concentrated nitric acid and hydrogen peroxide is ineffective for the dissolution of particulate lead compounds present in the test atmosphere (e.g. lead silicates), and an alternative, more vigorous dissolution procedure has not been used (e.g. employing hydrofluoric acid), then the analytical results will be subject to a negative bias. If it is desired to determine lead in samples containing high concentrations of silicates, consider the use of hydrofluoric acid in the digestion procedure.

12.2.3 Microwave digestion and flame atomic absorption spectrometry

Interlaboratory evaluations of lead determinations in reference materials (paint, soil and dust) have shown that microwave digestion with concentrated nitric acid performs equivalently to hot-plate digestion when followed by lead analysis using flame atomic absorption spectrometry [12].

12.2.4 Ultrasonic extraction and flame atomic absorption spectrometry

Ultrasonic extraction with electrochemical determination of lead has been shown to perform equivalently to microwave digestion and flame atomic absorption spectrometry in the determination of lead from laboratory-generated lead fume atmospheres [2].

12.2.5 Hot-plate digestion and electrothermal atomic absorption spectrometry

The detection limit of electrothermal atomic absorption measurements depends in part on the instrument used. However, the detection limit of the method has been estimated to be 0,003 μg per sample, and the precision of the measurement procedure was $< 0,05$ for samples in the range 0,1 μg to 4,5 μg . No bias has been identified. The applicable range is 0,01 μg to 0,5 μg Pb per sample [4], without dilution.

12.2.6 Microwave digestion and electrothermal atomic absorption spectrometry

Interlaboratory analysis of lead in reference materials has demonstrated the equivalence of microwave digestion followed by electrothermal atomic absorption spectrometry to hot-plate digestion followed by electrothermal atomic absorption spectrometric determination of lead [12].

12.2.7 Ultrasonic extraction and electrothermal atomic absorption spectrometry

Ultrasonic extraction with electrochemical detection of lead was evaluated against hot-plate extraction and electrothermal atomic absorption spectrometry for air samples collected from construction sites [13]. The test procedure using ultrasonic extraction was found to be equivalent to that using hot-plate digestion.

13 Special cases

13.1 If there is any doubt as to the suitability of the digestion or extraction procedure used for the dissolution of particulate lead compounds (e.g. silicates) that may be present in the test atmosphere, determine its effectiveness by analysing a bulk sample of known lead content that is similar in nature to the material being used or produced. If the efficiency of recovery is less than 90 %, use an alternative, more vigorous dissolution procedure, e.g. by using hydrofluoric acid. Do not use a correction factor to compensate for an apparently ineffective dissolution procedure.

It should be recognized that the recovery of lead can be dependent upon the particle size distribution of a bulk sample.

13.2 Anions that give rise to precipitates can interfere with lead analysis. If such interferents are likely to be present in sample solutions, add the disodium salt of ethylenediamine tetraacetic acid (EDTA) to the sample and blank solutions and to the calibration standard solutions, such that these solutions have a concentration of 0,1 mol/l of EDTA.

The addition of EDTA usually prevents precipitation, but high levels of phosphate can diminish the lead signal even in the presence of EDTA. If high levels of phosphate are suspected in the sample solutions, then the method of standard addition should be used to obtain accurate results (see 6.1.3 of ISO 6955:1982).

13.3 It has been postulated [14] and [15] that gaseous lead can be present in significant concentrations in certain work environments, e.g. when high temperature processes are used. In such circumstances, the sampling method described in this International Standard might not be fully effective because gaseous lead could pass through the filter. If necessary, this possibility can be investigated using a sampling train consisting of a filter and bubbler (see [14,15]).

13.4 When the filter transport cassettes or samplers are opened, it is advisable to look for evidence that particles have become dislodged from the filter during transportation. If this appears to have occurred, consider whether to discard the sample as invalid, or whether to wash the internal surfaces of the filter transport cassette or sampler into the sample dissolution vessel in order to recover the material concerned.

14 Test report

The test report shall contain the following information:

- statement to indicate the confidentiality of the information supplied, if appropriate;
- complete identification of the air sample, including the date of sampling, the place of sampling, the type of sample (personal or static), either the identity of the individual whose breathing zone was sampled (or other personal identifier) or the location at which the general occupational environment was sampled (for a static sample), a brief description of the work activities that were carried out during the sampling period, and a unique sample identification code;
- reference to this International Standard;
- make, type and diameter of filter used;
- make and type of sampler used;
- make and type of sampling pump used, and its identification;
- make and type of flowmeter used, the primary standard against which the calibration of the flowmeter was checked, the range of flowrates over which the calibration of the flowmeter was checked, and the atmospheric temperature and pressure at which the calibration of the flowmeter was checked, if appropriate (see 10.1.3);

- h) time at the start and at the end of the sampling period, and the duration of the sampling period, in minutes;
- i) mean flowrate during the sampling period, in litres per minute;
- j) mean atmospheric temperature and pressure during the sampling period, if appropriate (see 10.1.3);
- k) volume of air sampled, in litres, at ambient conditions;
- l) name of the person who collected the sample;
- m) time-weighted average mass concentration of lead found in the air sample (in mg/m³), at ambient temperature and pressure, or, if appropriate, adjusted to reference conditions;
- n) analytical variables used to calculate the result, including the concentrations of lead in the sample and blank solutions, the volumes of the sample and blank solutions, and the dilution factor, if applicable;

NOTE If the necessary data (e.g. the volume of air sampled) are not available to the laboratory for the above calculations to be carried out, the laboratory report may contain the analytical result in micrograms of lead per sample.

- o) type(s) of instrument(s) used for sample preparation and analysis, and unique identifiers(s);
- p) estimated detection limit under the working analytical conditions;
- q) any operation not specified in this International Standard, or regarded as optional;
- r) name of the analyst(s) [or other unique identifier(s)];
- s) date of the analysis; and
- t) any inadvertent deviations, unusual occurrences, or other notable observations.